
Midmark M11 - v1.0.5
Total Cycles: 4359

Name: RITTER2
Sterilizer ID: V2313228
02/19/2025 04:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.1 F 0.0
2:00 94.7 F 0.7
4:00 102.3 F 2.2
6:00 131.6 F 5.0
8:00 165.4 F 8.2
10:00 193.1 F 12.1
12:00 207.8 F 4.8
14:00 219.2 F 9.2
16:00 235.2 F 11.1
18:00 235.1 F 10.8
20:00 234.3 F 8.6
22:00 248.2 F 15.1
24:00 259.4 F 21.5
26:00 270.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.0
1:00 272.2 F 29.9
2:00 272.1 F 30.0
3:00 272.7 F 30.2
4:00 272.3 F 30.4

Min 271.9 F 29.9
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 204.0 F 0.0
10:00 185.1 F 0.0
15:00 185.5 F 0.0
20:00 183.2 F 0.0
25:00 183.5 F 0.0
30:00 184.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 27:36
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:06:12
CYCLE COMPLETE

02/19/2025 05:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____