
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006168
9:28 25/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:14
STERILIZING 4:17
136.0 C 220kPa 4:17
136.0 C 221kPa 4:47
136.3 C 222kPa 5:17
136.4 C 218kPa 5:47
136.4 C 224kPa 6:17
136.4 C 222kPa 6:47
136.2 C 221kPa 7:17
135.9 C 219kPa 7:47
135.9 C 219kPa 7:47
VENTING 7:48
AIR DRYING 8:35
CYCLE COMPLETE 38:35

Digital Signature #
52769FCFE2F7A5A5
