
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 000603
11:11 26/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:49
STERILIZING 4:00
136.0 C 229kPa 4:00
136.4 C 233kPa 4:30
136.4 C 233kPa 5:00
136.3 C 232kPa 5:30
136.3 C 232kPa 6:00
136.3 C 233kPa 6:30
136.4 C 233kPa 7:00
136.3 C 232kPa 7:30
136.3 C 232kPa 7:30
VENTING 7:31
AIR DRYING 8:25
CYCLE COMPLETE 10:17

Digital Signature #
ACD59A07B9864657
