

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3884

Name: RITTER2  
Sterilizer ID: V2313228  
08/22/2024 01:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.1 F 0.0  
2:00 99.2 F 0.0  
4:00 110.4 F 0.1  
6:00 149.5 F 1.7  
8:00 167.4 F 4.7  
10:00 180.1 F 8.1  
12:00 196.7 F 11.8  
14:00 215.3 F 5.1  
16:00 223.0 F 9.3  
18:00 232.3 F 11.9  
20:00 238.2 F 10.8  
22:00 235.7 F 8.9  
24:00 248.9 F 15.1  
26:00 259.8 F 21.1  
28:00 269.5 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 271.9 F 29.0  
2:00 272.1 F 29.2  
3:00 272.3 F 29.5  
4:00 271.4 F 29.3

Min 271.2 F 28.9  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.5 F 0.0  
5:00 210.5 F 0.0  
10:00 195.5 F 0.0  
15:00 186.4 F 0.0  
20:00 184.4 F 0.0  
25:00 182.3 F 0.0  
30:00 182.1 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 28:56  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:07:35  
CYCLE COMPLETE

08/22/2024 02:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_