
Midmark M11 - v1.0.0
Total Cycles: 4867

Name: RITTER 1
Sterilizer ID: V2178198
09/04/2024 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.1 F 0.0
2:00 109.5 F 0.0
4:00 142.7 F 0.4
6:00 173.3 F 0.7
8:00 196.7 F 3.2
10:00 209.3 F 7.2
12:00 221.7 F 11.2
14:00 224.0 F 5.3
16:00 237.0 F 10.5
18:00 234.3 F 7.6
20:00 246.7 F 13.5
22:00 244.3 F 12.0
24:00 257.0 F 18.7
26:00 268.3 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.8
1:00 272.4 F 28.2
2:00 272.5 F 28.6
3:00 271.6 F 27.7
4:00 272.8 F 28.5

Min 271.5 F 27.7
Max 272.9 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.9 F 0.0
5:00 199.6 F 0.0
10:00 201.1 F 0.0
15:00 203.0 F 0.0
20:00 207.3 F 0.0
25:00 206.2 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 27:11
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:05:06
CYCLE COMPLETE

09/04/2024 02:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____