
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006730
12:53 12/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:01
STERILIZING 3:05
136.5 C 221kPa 3:05
136.2 C 221kPa 3:35
136.4 C 221kPa 4:05
136.4 C 222kPa 4:35
136.5 C 223kPa 5:05
136.4 C 224kPa 5:35
136.4 C 225kPa 6:05
135.9 C 218kPa 6:35
135.9 C 218kPa 6:35
VENTING 6:35
AIR DRYING 7:27
CYCLE COMPLETE 7:46

Digital Signature #
86D3D90A48CC0F29
