

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1657

Name: RITTER2  
Sterilizer ID: V2313228  
10/26/2022 07:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.3 F 0.0  
2:00 76.4 F 0.0  
4:00 98.2 F 0.0  
6:00 111.0 F 0.0  
8:00 138.0 F 0.1  
10:00 175.1 F 0.2  
12:00 194.8 F 2.6  
14:00 207.0 F 5.7  
16:00 217.0 F 8.9  
18:00 226.9 F 12.4  
20:00 229.8 F 7.7  
22:00 240.0 F 12.4  
24:00 236.8 F 9.7  
26:00 237.6 F 9.3  
28:00 246.6 F 13.9  
30:00 257.6 F 19.8  
32:00 267.5 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.6 F 29.6  
2:00 271.7 F 28.7  
3:00 272.1 F 29.5  
4:00 272.2 F 29.2  
  
Min 271.5 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
5:00 202.8 F 0.0  
10:00 211.8 F 0.0  
15:00 214.1 F 0.0  
20:00 218.1 F 0.0  
25:00 218.1 F 0.0  
30:00 227.3 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 33:33  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:11:24  
CYCLE COMPLETE

10/26/2022 08:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_