
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004517
8:20 12/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:34
STERILIZING 3:37
136.5 C 221kPa 3:37
135.8 C 220kPa 4:07
136.4 C 222kPa 4:37
136.4 C 220kPa 5:07
136.3 C 222kPa 5:37
136.3 C 223kPa 6:07
136.4 C 223kPa 6:37
136.1 C 217kPa 7:07
136.1 C 217kPa 7:07
VENTING 7:07
AIR DRYING 7:56
CYCLE COMPLETE 8:02

Digital Signature #
DDA1F5CA4C6C5DB5
