
Midmark M11 - v1.0.5
Total Cycles: 566

Name: RITTER2
Sterilizer ID: V2313228
11/18/2021 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.0 F 0.0
2:00 162.9 F 3.2
4:00 181.5 F 7.9
6:00 204.4 F 13.5
8:00 219.5 F 8.6
10:00 232.0 F 10.7
12:00 239.6 F 12.1
14:00 239.8 F 11.0
16:00 253.2 F 17.7
18:00 264.5 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.4 F 30.0
2:00 271.7 F 29.3
3:00 271.4 F 29.7
4:00 272.1 F 29.9

Min 271.1 F 29.1
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 210.4 F 0.0
10:00 221.2 F 0.0
15:00 224.6 F 0.0
20:00 227.4 F 0.0
25:00 229.2 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

11/18/2021 02:18 PM
ACCEPTED: X
REJECTED:
Operator: U02