
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002943
11:24 12/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:50
STERILIZING 2:56
136.4 C 220kPa 2:56
136.4 C 225kPa 3:26
136.4 C 223kPa 3:56
136.4 C 222kPa 4:26
136.3 C 223kPa 4:56
136.1 C 219kPa 5:26
136.4 C 221kPa 5:56
136.4 C 225kPa 6:26
136.4 C 225kPa 6:26
VENTING 6:26
AIR DRYING 7:19
CYCLE COMPLETE 7:21

Digital Signature #
6D548AD180AD415F
