
Midmark M11 - v1.0.5
Total Cycles: 4223

Name: RITTER2
Sterilizer ID: V2313228
12/17/2024 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.0 F 0.0
2:00 152.3 F 2.3
4:00 160.9 F 5.8
6:00 191.2 F 10.4
8:00 209.7 F 4.6
10:00 225.2 F 9.6
12:00 230.6 F 8.8
14:00 239.9 F 12.4
16:00 239.9 F 11.1
18:00 253.3 F 17.8
20:00 264.8 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.3
1:00 271.9 F 29.8
2:00 272.5 F 29.6
3:00 271.4 F 29.3
4:00 272.5 F 30.1

Min 271.3 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 208.1 F 0.0
10:00 207.4 F 0.0
15:00 199.6 F 0.0
20:00 194.8 F 0.0
25:00 192.0 F 0.0
30:00 191.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

12/17/2024 04:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____