

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4208

Name: RITTER2  
Sterilizer ID: V2313228  
12/12/2024 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 89.3 F 0.0  
2:00 89.8 F 0.0  
4:00 96.6 F 0.0  
6:00 129.1 F 1.0  
8:00 156.5 F 3.8  
10:00 173.0 F 7.0  
12:00 192.3 F 10.2  
14:00 208.9 F 10.3  
16:00 218.9 F 7.9  
18:00 228.2 F 12.1  
20:00 232.0 F 8.8  
22:00 242.1 F 12.1  
24:00 242.5 F 12.3  
26:00 253.5 F 18.0  
28:00 263.6 F 24.0  
30:00 272.6 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.9  
1:00 272.6 F 30.0  
2:00 272.7 F 30.4  
3:00 271.9 F 30.1  
4:00 272.5 F 30.1

Min 271.9 F 29.8  
Max 272.9 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
5:00 230.9 F 0.0  
10:00 239.1 F 0.0  
15:00 218.2 F 0.0  
20:00 196.0 F 0.0  
25:00 185.1 F 0.0  
30:00 185.2 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 31:03  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:09:48  
CYCLE COMPLETE

12/12/2024 03:52 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_