
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001225
14:18 19/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:49
PRESSURIZING 3:53
STERILIZING 5:13
136.0 C 220kPa 5:13
136.1 C 222kPa 5:43
136.3 C 224kPa 6:13
136.4 C 226kPa 6:43
136.3 C 224kPa 7:13
136.3 C 224kPa 7:43
136.4 C 225kPa 8:13
136.3 C 224kPa 8:43
136.3 C 224kPa 8:43
VENTING 8:43
AIR DRYING 9:41
CYCLE COMPLETE 15:21

Digital Signature #
BEE5925A4C4CBAE1
