
Midmark M11 - v1.0.5
Total Cycles: 1247

Name: RITTER2
Sterilizer ID: V2313228
07/04/2022 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.3 F 0.0
2:00 122.4 F 1.0
4:00 132.7 F 3.3
6:00 158.3 F 6.7
8:00 180.6 F 10.6
10:00 213.3 F 4.6
12:00 220.8 F 8.2
14:00 232.2 F 12.8
16:00 234.4 F 9.4
18:00 237.4 F 9.8
20:00 245.3 F 13.3
22:00 256.8 F 19.5
24:00 267.1 F 26.0
26:00 273.0 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.6
1:00 272.4 F 29.5
2:00 272.6 F 29.8
3:00 272.6 F 30.1
4:00 271.7 F 29.8

Min 271.6 F 29.5
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 226.8 F 0.0
10:00 227.1 F 0.0
15:00 219.4 F 0.0
20:00 216.4 F 0.0
25:00 218.0 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:13
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:04:17
CYCLE COMPLETE

07/04/2022 10:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____