
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000710
14:12 28/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:49
STERILIZING 3:54
136.0 C 220kPa 3:54
136.4 C 225kPa 4:24
136.3 C 224kPa 4:54
136.4 C 224kPa 5:24
136.1 C 223kPa 5:54
136.4 C 224kPa 6:24
136.2 C 220kPa 6:54
136.4 C 224kPa 7:24
136.4 C 224kPa 7:24
VENTING 7:24
AIR DRYING 8:17
CYCLE COMPLETE 13:23

Digital Signature #
0CADCF0BC02B7B44
