
Midmark M11 - v1.0.5
Total Cycles: 2362

Name: RITTER2
Sterilizer ID: V2313228
05/12/2023 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 181.5 F 3.4
4:00 185.9 F 7.5
6:00 208.8 F 12.9
8:00 224.6 F 8.0
10:00 236.4 F 13.4
12:00 238.9 F 11.2
14:00 237.9 F 9.9
16:00 251.7 F 16.8
18:00 263.4 F 23.8
20:00 271.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 29.5
2:00 271.4 F 29.4
3:00 272.5 F 29.6
4:00 271.3 F 29.1

Min 271.3 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 227.1 F 0.0
10:00 228.3 F 0.0
15:00 225.0 F 0.0
20:00 223.5 F 0.0
25:00 223.1 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:02
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

05/12/2023 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____