

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005652  
9:11 02/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:06  
STERILIZING 3:13  
136.5 C 224kPa 3:13  
136.2 C 223kPa 3:43  
136.2 C 223kPa 4:13  
136.0 C 218kPa 4:43  
136.4 C 223kPa 5:13  
136.3 C 224kPa 5:43  
136.2 C 223kPa 6:13  
136.3 C 223kPa 6:43  
136.3 C 223kPa 6:43  
VENTING 6:43  
AIR DRYING 7:35  
CYCLE COMPLETE 11:42

-----  
Digital Signature #  
FD9B7EAD6B4CCBC4  
-----