
Midmark M11 - v1.0.5
Total Cycles: 3948

Name: RITTER2
Sterilizer ID: V2313228
09/16/2024 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.3 F 0.0
2:00 98.5 F 0.0
4:00 105.4 F 0.0
6:00 143.0 F 0.1
8:00 172.4 F 2.3
10:00 193.0 F 5.3
12:00 200.6 F 8.9
14:00 211.5 F 12.6
16:00 221.5 F 6.7
18:00 232.3 F 10.9
20:00 232.0 F 8.0
22:00 242.9 F 13.0
24:00 241.9 F 11.6
26:00 253.2 F 17.4
28:00 263.3 F 23.3
30:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 271.6 F 29.1
2:00 272.6 F 29.6
3:00 272.1 F 29.1
4:00 271.3 F 28.8

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 193.4 F 0.0
10:00 205.2 F 0.0
15:00 196.7 F 0.0
20:00 200.6 F 0.0
25:00 203.9 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:18
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:08:52
CYCLE COMPLETE

09/16/2024 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____