
Midmark M11 - v1.0.0
Total Cycles: 4334

Name: RITTER 1
Sterilizer ID: V2178198
02/26/2024 03:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.2 F 0.0
2:00 92.1 F 0.6
4:00 100.8 F 2.2
6:00 130.7 F 4.6
8:00 175.7 F 7.5
10:00 191.5 F 11.2
12:00 211.6 F 3.5
14:00 210.3 F 5.6
16:00 221.2 F 9.1
18:00 231.5 F 13.0
20:00 233.0 F 9.0
22:00 243.1 F 11.9
24:00 242.4 F 11.7
26:00 253.2 F 17.1
28:00 262.5 F 22.9
30:00 272.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.8 F 29.1
2:00 272.1 F 28.9
3:00 271.1 F 28.6
4:00 272.3 F 29.3

Min 271.0 F 28.4
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 232.4 F 0.0
10:00 211.7 F 0.0
15:00 205.7 F 0.0
20:00 220.1 F 0.0
25:00 229.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 30:21
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:09:03
CYCLE COMPLETE

02/26/2024 04:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____