
Midmark M11 - v1.0.5
Total Cycles: 4195

Name: RITTER2
Sterilizer ID: V2313228
12/06/2024 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.4 F 0.0
2:00 106.0 F 0.0
4:00 127.4 F 0.1
6:00 161.1 F 0.9
8:00 187.5 F 3.5
10:00 203.3 F 6.9
12:00 216.3 F 10.5
14:00 229.9 F 7.8
16:00 233.4 F 9.1
18:00 237.7 F 13.5
20:00 239.6 F 11.0
22:00 237.3 F 9.5
24:00 250.4 F 15.7
26:00 261.2 F 21.9
28:00 271.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.8 F 29.4
2:00 272.6 F 29.9
3:00 271.7 F 29.1
4:00 272.7 F 29.7

Min 271.5 F 29.1
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.4 F 0.0
5:00 230.3 F 0.0
10:00 223.4 F 0.0
15:00 182.1 F 0.0
20:00 177.9 F 0.0
25:00 175.0 F 0.0
30:00 175.2 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 29:26
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:08:17
CYCLE COMPLETE

12/06/2024 05:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____