

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3369

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2023 02:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
2:00 191.5 F 0.9  
4:00 203.8 F 6.1  
6:00 219.5 F 11.1  
8:00 224.3 F 6.3  
10:00 237.6 F 11.7  
12:00 236.7 F 9.4  
14:00 238.2 F 9.2  
16:00 249.5 F 14.9  
18:00 261.7 F 21.7  
20:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.4 F 28.5  
2:00 272.0 F 28.6  
3:00 272.1 F 28.4  
4:00 272.1 F 28.7  
  
Min 271.4 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 216.1 F 0.0  
10:00 233.7 F 0.0  
15:00 230.0 F 0.0  
20:00 221.3 F 0.0  
25:00 218.3 F 0.0  
30:00 216.5 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 20:22  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:34  
CYCLE COMPLETE

06/09/2023 03:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_