

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001312  
14:43 20/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:53  
PRESSURIZING 3:55  
STERILIZING 5:33  
136.4 C 224kPa 5:33  
136.4 C 226kPa 6:03  
136.2 C 222kPa 6:33  
136.2 C 223kPa 7:03  
136.3 C 224kPa 7:33  
136.2 C 224kPa 8:03  
136.4 C 224kPa 8:33  
136.4 C 223kPa 9:03  
136.4 C 223kPa 9:03  
VENTING 9:03  
AIR DRYING 10:01  
CYCLE COMPLETE 25:02  
-----

Digital Signature #  
83B71FBEBA7DFB7D  
-----