

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2449

Name: RITTER2  
Sterilizer ID: V2313228  
06/01/2023 08:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.6 F 0.0  
2:00 189.7 F 2.5  
4:00 184.8 F 6.7  
6:00 208.5 F 12.2  
8:00 223.3 F 7.9  
10:00 238.2 F 12.4  
12:00 239.8 F 11.6  
14:00 239.2 F 10.4  
16:00 252.7 F 17.3  
18:00 264.6 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.6 F 29.8  
2:00 271.6 F 29.1  
3:00 271.8 F 29.7  
4:00 273.1 F 30.1  
  
Min 271.0 F 28.9  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.0 F 0.0  
5:00 216.8 F 0.0  
10:00 234.9 F 0.0  
15:00 229.6 F 0.0  
20:00 223.1 F 0.0  
25:00 222.1 F 0.0  
30:00 223.1 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 19:53  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:49  
CYCLE COMPLETE

06/01/2023 09:26 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_