
Midmark M11 - v1.0.5
Total Cycles: 302

Name: RITTER2
Sterilizer ID: V2313228
08/23/2021 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	102.2 F	0.0
2:00	104.4 F	1.0
4:00	134.0 F	3.8
6:00	163.0 F	7.4
8:00	183.1 F	12.0
10:00	217.8 F	6.2
12:00	228.9 F	11.1
14:00	226.8 F	7.1
16:00	242.0 F	12.8
18:00	239.7 F	11.1
20:00	252.3 F	17.3
22:00	263.6 F	23.8
24:00	272.0 F	29.4

STERILIZING

mm:ss	Degrees	PSI
0:00	271.6 F	29.7
1:00	271.8 F	29.9
2:00	271.6 F	29.9
3:00	271.5 F	29.8
4:00	271.9 F	29.6

Min 271.5 F 29.6
Max 273.0 F 30.3

VENTING CHAMBER

mm:ss	Degrees	PSI
0:00	184.1 F	0.0
5:00	229.9 F	0.0
10:00	219.1 F	0.0
15:00	225.5 F	0.0
20:00	214.2 F	0.0
25:00	214.3 F	0.0
30:00	210.8 F	0.0

Drying Complete

FILLING: 1:10
HEATING: 24:47
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:03:01
CYCLE COMPLETE

08/23/2021 01:34 PM
ACCEPTED: X
REJECTED:
Operator: U02