

-----  
Midmark M11 - v1.0.0  
Total Cycles: 789

Name: RITTER 1  
Sterilizer ID: V2178198  
04/21/2021 06:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.6 F 0.0  
2:00 77.0 F 0.0  
4:00 81.8 F 0.0  
6:00 113.4 F 0.1  
8:00 150.5 F 0.2  
10:00 178.6 F 0.3  
12:00 198.7 F 2.5  
14:00 205.5 F 5.4  
16:00 216.3 F 8.4  
18:00 227.8 F 11.8  
20:00 226.5 F 7.0  
22:00 239.1 F 12.1  
24:00 236.9 F 9.8  
26:00 237.1 F 8.7  
28:00 249.0 F 15.1  
30:00 260.9 F 21.8  
32:00 271.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.3 F 28.8  
2:00 273.0 F 29.4  
3:00 271.8 F 28.8  
4:00 272.4 F 29.0  
  
Min 271.5 F 28.5  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
5:00 209.0 F 0.0  
10:00 229.5 F 0.0  
15:00 225.8 F 0.0  
20:00 222.4 F 0.0  
25:00 182.6 F 0.0  
30:00 214.3 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 32:31  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:10:30  
CYCLE COMPLETE

04/21/2021 08:08 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01