
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001545
11:42 10/29/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:54
PRESSURIZING 6:37
STERILIZING 8:08
136.0 C 220kPa 8:08
136.4 C 219kPa 8:38
136.4 C 223kPa 9:08
136.4 C 224kPa 9:38
136.0 C 221kPa 10:08
136.4 C 223kPa 10:38
136.4 C 223kPa 11:08
136.3 C 221kPa 11:38
VENTING 11:38
AIR DRYING 12:41
CYCLE COMPLETE 42:41

Digital Signature #
037D6FACA95EB CD2
