
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001653
13:48 06/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:16
STERILIZING 3:20
136.0 C 222kPa 3:20
136.3 C 224kPa 3:50
136.4 C 225kPa 4:20
136.4 C 226kPa 4:50
136.4 C 225kPa 5:20
136.4 C 224kPa 5:50
136.4 C 225kPa 6:20
136.2 C 224kPa 6:50
136.2 C 224kPa 6:50
VENTING 6:50
AIR DRYING 7:43
CYCLE COMPLETE 7:47

Digital Signature #
FFC9E1786AF58816
