
Midmark M11 - v1.0.5
Total Cycles: 4446

Name: RITTER2
Sterilizer ID: V2313228
03/24/2025 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.2 F 0.0
2:00 175.0 F 1.5
4:00 173.1 F 4.9
6:00 188.7 F 9.2
8:00 205.7 F 10.9
10:00 223.3 F 7.9
12:00 237.6 F 12.3
14:00 234.5 F 9.5
16:00 239.5 F 10.0
18:00 245.5 F 13.9
20:00 256.9 F 20.0
22:00 267.2 F 26.4
24:00 273.4 F 30.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.1
1:00 271.9 F 30.1
2:00 272.0 F 30.0
3:00 272.1 F 30.1
4:00 272.1 F 30.3

Min 271.6 F 30.0
Max 272.6 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 242.6 F 0.0
10:00 226.0 F 0.0
15:00 225.6 F 0.0
20:00 216.8 F 0.0
25:00 213.8 F 0.0
30:00 208.8 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 24:16
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:03:32
CYCLE COMPLETE

03/24/2025 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____