
Midmark M11 - v1.0.0
Total Cycles: 4394

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2024 04:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.6 F 0.0
2:00 152.2 F 2.6
4:00 165.0 F 6.3
6:00 196.0 F 10.4
8:00 213.5 F 3.9
10:00 220.8 F 7.9
12:00 232.7 F 12.6
14:00 235.1 F 9.2
16:00 240.0 F 10.3
18:00 246.0 F 13.2
20:00 257.7 F 19.5
22:00 268.4 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.0
2:00 271.4 F 28.2
3:00 272.5 F 29.1
4:00 272.0 F 28.5

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 212.1 F 0.0
10:00 217.5 F 0.0
15:00 219.4 F 0.0
20:00 229.7 F 0.0
25:00 237.2 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 23:15
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:05
CYCLE COMPLETE

03/21/2024 05:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____