
Midmark M11 - v1.0.0
Total Cycles: 4727

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2024 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.2 F 0.0
2:00 178.4 F 2.3
4:00 181.8 F 5.9
6:00 195.4 F 10.0
8:00 217.5 F 7.1
10:00 220.3 F 7.2
12:00 231.4 F 11.4
14:00 230.7 F 7.7
16:00 243.3 F 13.2
18:00 242.0 F 11.5
20:00 253.5 F 17.5
22:00 264.2 F 23.7
24:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.5
2:00 271.8 F 28.8
3:00 272.0 F 29.3
4:00 272.3 F 29.3

Min 271.2 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 209.1 F 0.0
10:00 211.8 F 0.0
15:00 210.5 F 0.0
20:00 224.7 F 0.0
25:00 232.9 F 0.0
30:00 234.4 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 24:13
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:03:23
CYCLE COMPLETE

07/15/2024 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____