
Midmark M11 - v1.0.5
Total Cycles: 2912

Name: RITTER2
Sterilizer ID: V2313228
09/20/2023 02:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.3 F 0.0
2:00 169.7 F 3.3
4:00 190.2 F 7.8
6:00 206.8 F 12.8
8:00 221.9 F 7.6
10:00 234.0 F 12.9
12:00 236.6 F 10.6
14:00 236.0 F 9.3
16:00 250.0 F 16.0
18:00 261.5 F 22.6
20:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.4 F 29.6
2:00 271.3 F 29.5
3:00 272.1 F 29.6
4:00 271.1 F 29.4

Min 271.1 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 206.7 F 0.0
10:00 209.5 F 0.0
15:00 213.7 F 0.0
20:00 209.6 F 0.0
25:00 215.6 F 0.0
30:00 211.0 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

09/20/2023 03:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____