
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004115
15:03 26/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:14
STERILIZING 3:11
136.0 C 220kPa 3:11
136.4 C 222kPa 3:41
136.2 C 224kPa 4:11
136.4 C 225kPa 4:41
136.4 C 225kPa 5:11
136.1 C 223kPa 5:41
136.2 C 223kPa 6:11
136.2 C 224kPa 6:41
136.2 C 224kPa 6:41
VENTING 6:41
AIR DRYING 7:28
CYCLE COMPLETE 7:33

Digital Signature #
8C1CA2D7CD41DD03
