
Midmark M11 - v1.0.5
Total Cycles: 4111

Name: RITTER2
Sterilizer ID: V2313228
11/13/2024 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 82.7 F 0.0
4:00 84.9 F 0.1
6:00 123.4 F 0.6
8:00 162.1 F 0.6
10:00 188.1 F 1.4
12:00 200.9 F 4.8
14:00 213.0 F 8.5
16:00 224.5 F 12.4
18:00 229.0 F 7.4
20:00 240.0 F 12.2
22:00 238.2 F 9.8
24:00 238.3 F 9.0
26:00 250.1 F 15.2
28:00 261.8 F 21.9
30:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.5 F 28.6
2:00 272.4 F 29.0
3:00 271.8 F 28.2
4:00 272.8 F 29.0

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 221.7 F 0.0
10:00 196.2 F 0.0
15:00 194.8 F 0.0
20:00 223.2 F 0.0
25:00 230.0 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:24
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:08:31
CYCLE COMPLETE

11/13/2024 10:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____