

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5336

Name: RITTER 1  
Sterilizer ID: V2178198  
02/12/2025 11:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.6 F 0.0  
2:00 119.2 F 1.2  
4:00 136.5 F 3.9  
6:00 162.9 F 6.8  
8:00 183.9 F 10.0  
10:00 198.7 F 10.6  
12:00 212.9 F 5.5  
14:00 232.7 F 9.4  
16:00 237.0 F 10.6  
18:00 238.9 F 9.8  
20:00 238.9 F 9.3  
22:00 246.3 F 13.0  
24:00 256.5 F 18.8  
26:00 266.9 F 25.0  
28:00 272.2 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.5  
1:00 272.7 F 29.0  
2:00 271.6 F 28.6  
3:00 272.5 F 28.7  
4:00 272.2 F 29.0

Min 271.2 F 28.4  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 217.3 F 0.0  
10:00 249.1 F 0.0  
15:00 223.9 F 0.0  
20:00 225.5 F 0.0  
25:00 225.3 F 0.0  
30:00 225.9 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 28:20  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:06:54  
CYCLE COMPLETE

02/12/2025 12:24 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_