

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001480  
15:15 25/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:37  
STERILIZING 3:42  
136.0 C 222kPa 3:42  
136.4 C 226kPa 4:12  
136.2 C 223kPa 4:42  
136.4 C 227kPa 5:12  
136.4 C 226kPa 5:42  
136.3 C 226kPa 6:12  
136.4 C 225kPa 6:42  
136.4 C 226kPa 7:12  
136.4 C 226kPa 7:12  
VENTING 7:12  
AIR DRYING 8:07  
CYCLE COMPLETE 9:10  
-----

Digital Signature #  
3D0CC37D3317DEC0  
-----