
Midmark M11 - v1.0.0
Total Cycles: 4446

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2024 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 155.5 F 1.5
4:00 156.3 F 3.8
6:00 167.6 F 7.1
8:00 201.1 F 11.3
10:00 213.2 F 4.4
12:00 225.6 F 8.6
14:00 236.5 F 13.1
16:00 233.7 F 9.5
18:00 240.2 F 10.1
20:00 245.0 F 13.0
22:00 256.3 F 18.9
24:00 266.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.2 F 29.3
2:00 271.8 F 28.8
3:00 271.0 F 28.6
4:00 272.3 F 29.6

Min 271.0 F 28.4
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 222.5 F 0.0
10:00 228.0 F 0.0
15:00 226.7 F 0.0
20:00 230.1 F 0.0
25:00 229.3 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:42
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:04:27
CYCLE COMPLETE

04/11/2024 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____