
Midmark M11 - v1.0.5
Total Cycles: 4266

Name: RITTER2
Sterilizer ID: V2313228
01/15/2025 12:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.3 F 0.0
2:00 109.6 F 0.0
4:00 112.5 F 0.0
6:00 141.6 F 0.1
8:00 176.8 F 0.4
10:00 196.1 F 1.7
12:00 213.4 F 5.1
14:00 224.0 F 8.9
16:00 237.0 F 12.9
18:00 235.5 F 9.6
20:00 239.4 F 10.3
22:00 244.8 F 13.0
24:00 243.3 F 12.0
26:00 254.9 F 18.1
28:00 265.3 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.4
2:00 272.2 F 28.8
3:00 271.3 F 28.8
4:00 272.5 F 29.4

Min 271.2 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 231.3 F 0.0
10:00 218.1 F 0.0
15:00 236.8 F 0.0
20:00 233.4 F 0.0
25:00 230.4 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:59
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:08:36
CYCLE COMPLETE

01/15/2025 01:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____