
Midmark M11 - v1.0.0
Total Cycles: 2936

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2023 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.5 F 0.0
2:00 193.4 F 2.8
4:00 197.7 F 7.6
6:00 216.4 F 13.0
8:00 227.3 F 8.0
10:00 238.7 F 12.7
12:00 239.4 F 11.0
14:00 237.3 F 9.2
16:00 251.7 F 16.1
18:00 263.7 F 23.1
20:00 271.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.2
2:00 272.3 F 28.7
3:00 271.3 F 28.2
4:00 272.2 F 29.1

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 218.8 F 0.0
10:00 229.5 F 0.0
15:00 239.6 F 0.0
20:00 230.3 F 0.0
25:00 213.1 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

02/23/2023 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____