

-----  
Midmark M11 - v1.0.5  
Total Cycles: 907

Name: RITTER2  
Sterilizer ID: V2313228  
03/25/2022 10:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
2:00 183.8 F 1.7  
4:00 200.9 F 6.7  
6:00 217.1 F 12.4  
8:00 231.7 F 8.8  
10:00 236.7 F 10.4  
12:00 244.4 F 13.3  
14:00 244.3 F 12.8  
16:00 257.5 F 19.8  
18:00 269.0 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 29.3  
1:00 272.7 F 29.9  
2:00 271.5 F 29.4  
3:00 272.8 F 29.7  
4:00 271.8 F 29.6  
  
Min 271.3 F 29.2  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.9 F 0.0  
5:00 235.3 F 0.0  
10:00 224.4 F 0.0  
15:00 230.1 F 0.0  
20:00 224.1 F 0.0  
25:00 229.0 F 0.0  
30:00 227.3 F 0.0  
  
Drying Complete

FILLING: 1:48  
HEATING: 19:41  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

03/25/2022 12:16 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02