
Midmark M11 - v1.0.5
Total Cycles: 3294

Name: RITTER2
Sterilizer ID: V2313228
01/12/2024 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.2 F 0.0
2:00 128.4 F 0.5
4:00 137.3 F 2.6
6:00 165.0 F 5.9
8:00 187.7 F 9.9
10:00 209.7 F 7.5
12:00 218.1 F 7.8
14:00 232.9 F 12.4
16:00 233.1 F 9.2
18:00 239.1 F 10.4
20:00 245.6 F 13.6
22:00 257.1 F 19.8
24:00 267.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.7
2:00 271.4 F 29.0
3:00 272.2 F 29.9
4:00 272.0 F 29.4

Min 271.2 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 206.4 F 0.0
10:00 208.7 F 0.0
15:00 198.2 F 0.0
20:00 194.3 F 0.0
25:00 193.0 F 0.0
30:00 192.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:24
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:55
CYCLE COMPLETE

01/12/2024 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____