
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000806
10:13 12/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:00
STERILIZING 3:03
136.5 C 223kPa 3:03
136.0 C 222kPa 3:33
136.4 C 221kPa 4:03
136.4 C 221kPa 4:33
136.2 C 223kPa 5:03
136.4 C 223kPa 5:33
136.4 C 222kPa 6:03
136.3 C 221kPa 6:33
136.3 C 221kPa 6:33
VENTING 6:33
AIR DRYING 7:27
CYCLE COMPLETE 8:01

Digital Signature #
2905169A2E382665
