
Midmark M11 - v1.0.5
Total Cycles: 4473

Name: RITTER2
Sterilizer ID: V2313228
03/31/2025 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.3 F 0.0
2:00 132.2 F 1.0
4:00 133.3 F 3.0
6:00 169.8 F 6.3
8:00 190.2 F 10.6
10:00 218.8 F 5.7
12:00 224.8 F 8.7
14:00 236.2 F 13.2
16:00 234.5 F 9.8
18:00 238.6 F 9.7
20:00 244.9 F 13.5
22:00 256.0 F 19.5
24:00 266.1 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.5
1:00 272.2 F 30.6
2:00 272.3 F 30.4
3:00 272.1 F 30.3
4:00 272.0 F 30.2

Min 270.5 F 29.4
Max 273.2 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 195.3 F 0.0
10:00 191.1 F 0.0
15:00 194.8 F 0.0
20:00 195.9 F 0.0
25:00 196.8 F 0.0
30:00 199.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:40
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:24
CYCLE COMPLETE

03/31/2025 12:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____