
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003578
15:00 02/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:29
STERILIZING 3:35
136.4 C 220kPa 3:35
136.3 C 223kPa 4:05
136.2 C 222kPa 4:35
136.4 C 224kPa 5:05
136.4 C 222kPa 5:35
136.4 C 222kPa 6:05
136.3 C 222kPa 6:35
136.3 C 222kPa 7:05
136.3 C 222kPa 7:05
VENTING 7:05
AIR DRYING 8:06
CYCLE COMPLETE 8:08

Digital Signature #
879F3724DB1EA7C3
