
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000055
12:49 19/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:18
STERILIZING 4:46
136.4 C 223kPa 4:46
136.3 C 224kPa 5:17
136.4 C 225kPa 5:47
136.3 C 223kPa 6:17
136.2 C 223kPa 6:47
136.4 C 225kPa 7:17
136.4 C 225kPa 7:47
136.4 C 224kPa 8:17
136.4 C 224kPa 8:17
VENTING 8:17
AIR DRYING 9:08
CYCLE COMPLETE 14:25

Digital Signature #
68364A2EDBC76C7E
