
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 003770
15:25 06/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:29
STERILIZING 3:36
136.4 C 224kPa 3:36
136.4 C 225kPa 4:06
136.4 C 224kPa 4:36
136.4 C 224kPa 5:06
136.2 C 220kPa 5:36
136.0 C 221kPa 6:06
136.4 C 225kPa 6:36
136.4 C 222kPa 7:06
136.4 C 222kPa 7:06
VENTING 7:07
AIR DRYING 8:02
CYCLE COMPLETE 8:45

Digital Signature #
07A259B293A66EE0
