
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001222
16:10 18/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:48
STERILIZING 4:00
136.3 C 221kPa 4:00
136.4 C 225kPa 4:30
136.2 C 223kPa 5:00
136.4 C 224kPa 5:30
136.1 C 222kPa 6:00
136.2 C 223kPa 6:30
136.4 C 225kPa 7:00
136.2 C 222kPa 7:30
136.2 C 222kPa 7:30
VENTING 7:30
AIR DRYING 8:21
CYCLE COMPLETE 23:21

Digital Signature #
56FA4F7E5A0C2DCD
