
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005839
10:03 26/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:03
STERILIZING 3:03
136.0 C 220kPa 3:03
136.4 C 225kPa 3:33
136.2 C 223kPa 4:03
136.0 C 220kPa 4:33
136.2 C 220kPa 5:03
136.4 C 225kPa 5:33
136.1 C 222kPa 6:03
136.3 C 221kPa 6:33
136.3 C 221kPa 6:33
VENTING 6:33
AIR DRYING 7:24
CYCLE COMPLETE 7:28

Digital Signature #
D337376D1CE0F393
