
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001977
11:20 26/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:38
STERILIZING 3:30
136.0 C 220kPa 3:30
136.4 C 225kPa 4:00
136.4 C 226kPa 4:30
136.4 C 223kPa 5:00
136.4 C 226kPa 5:30
136.3 C 226kPa 6:00
136.4 C 225kPa 6:30
136.4 C 226kPa 7:00
136.4 C 226kPa 7:00
VENTING 7:00
AIR DRYING 7:51
CYCLE COMPLETE 28:36

Digital Signature #
810C30168A05CC9D
