
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000132
10:09 12/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:41
STERILIZING 4:49
136.0 C 220kPa 4:49
135.8 C 220kPa 5:19
136.0 C 221kPa 5:49
136.3 C 219kPa 6:19
136.4 C 224kPa 6:49
136.4 C 224kPa 7:19
136.4 C 224kPa 7:49
136.4 C 222kPa 8:19
136.4 C 222kPa 8:19
VENTING 8:19
AIR DRYING 9:13
CYCLE COMPLETE 9:33

Digital Signature #
D5CF0987F265FE75
