

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 006143  
15:23 16/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:14  
PRESSURIZING 2:22  
STERILIZING 3:19  
136.4 C 224kPa 3:19  
136.3 C 224kPa 3:49  
136.3 C 224kPa 4:19  
136.4 C 223kPa 4:49  
136.2 C 222kPa 5:19  
136.3 C 223kPa 5:49  
136.4 C 225kPa 6:19  
136.2 C 221kPa 6:49  
136.2 C 221kPa 6:49  
VENTING 6:50  
AIR DRYING 7:39  
CYCLE COMPLETE 8:06

-----  
Digital Signature #  
96D1E98E1A197652  
-----